

UNITED STATES DISTRICT COURT  
DISTRICT OF NEW JERSEY

UNITED STATES OF AMERICA : Hon.  
v. : Crim. No. 18-  
OLYMPUS MEDICAL SYSTEMS : 21 U.S.C. §§ 331(a) and  
CORPORATION : 333(a)(1)

I N F O R M A T I O N

The Acting United States Attorney for the District of New Jersey charges:

BACKGROUND

At all times relevant to this Information, unless otherwise alleged:

Defendant Olympus Medical Systems Corporation  
and the TJF-Q180V Duodenoscope

1. Olympus Corporation is a multinational manufacturer of optical imaging, laboratory, and medical equipment. Olympus Corporation is headquartered in Tokyo, Japan, is listed on the Tokyo Stock Exchange, and has subsidiaries throughout the world, including in the United States.

2. OLYMPUS MEDICAL SYSTEMS CORPORATION ("OMSC") is a wholly owned subsidiary of Olympus Corporation, and is located

in Tokyo, Japan. OMSC developed and manufactured endoscopes, including duodenoscopes, for direct internal observations of the human body.

3. Duodenoscopes are flexible, lighted tubes that are threaded through the mouth, throat, and stomach into the top of the small intestine (duodenum). The end of the tube has a light, camera, and forceps elevator, which is controlled by an elevator wire that passes through a channel in the tube.

4. Duodenoscopes are used during endoscopic retrograde cholangiopancreatography ("ERCP"), a potentially life-saving procedure to diagnose and treat problems in the pancreas and bile ducts. Duodenoscopes are used throughout the world, including within the United States, where duodenoscopes are used in more than 500,000 ERCP procedures each year.

5. Because duodenoscopes are reusable devices, duodenoscopes must be reprocessed (cleaned) after each use by a procedure established by the manufacturer. If a reprocessing is unsuccessful, infectious material may remain on or in the duodenoscope, and subsequent patients treated with the duodenoscope may become infected, which may lead to serious illness or death.

6. Reprocessing a duodenoscope involves an initial "pre-cleaning" step in which a technician manually washes the duodenoscope with fluids and a brush and a second step that can

also be done manually but that most commonly is done automatically by placing the scope in a dishwasher-type machine, called an automated endoscope reprocessor.

7. Between August 2012 and October 2014, Olympus Corporation and its subsidiaries had approximately 85% of the United States market for duodenoscopes. During this time, global sales of all Olympus medical devices accounted for approximately 75% of Olympus Corporation's revenue.

8. In 2010, Olympus America Inc. ("OAI"), another wholly owned, indirect subsidiary of Olympus Corporation, began marketing and distributing the TJF-Q180V duodenoscope ("Q180V") in the United States. OMSC manufactured the Q180V.

9. Unlike previous Olympus duodenoscopes, the Q180V had a closed elevator wire channel. The Q180V's sealed channel was intended to prevent bodily fluids from entering the elevator wire channel, thus, according to OMSC, eliminating the need to clean the elevator wire channel.

#### FDA and the FDCA

10. The U.S. Food and Drug Administration ("FDA") is responsible for protecting the health and safety of the American public by assuring, among other things, that medical devices intended for use in the treatment of human beings are safe and effective for their intended uses. Pursuant to its statutory mandate, FDA regulates the manufacture, processing, packing,

labeling, and shipment in interstate commerce of medical devices.

11. The Federal Food, Drug, and Cosmetic Act ("FDCA"), among other things, governs the manufacture and interstate distribution of medical devices for human use, as codified at Title 21, United States Code, Sections 301-399f.

**Medical Device Reporting**

12. The FDCA and its implementing regulations provide a mechanism that allows FDA, and others, to identify and monitor adverse events (deaths and serious injuries) and certain malfunctions involving medical devices.

13. Pursuant to 21 U.S.C. § 360i(a) and 21 CFR Part 803, medical device manufacturers must (1) develop, maintain, and implement written procedures for the identification and evaluation of all malfunctions, serious injuries, and deaths to determine whether a Medical Device Report ("MDR") is required for an event; (2) submit MDR reportable events involving their medical devices to FDA; and (3) establish and maintain complete files for all MDR events. These requirements apply to all manufacturers of medical devices in the United States, including foreign manufacturers who export devices to the United States, such as OMSC.

14. Manufacturers must file an MDR with FDA within thirty (30) days of receiving or becoming aware of information

that reasonably suggests that a device the manufacturer markets

(a) may have caused or contributed to a death or serious injury

or (b) has malfunctioned and the device or a similar device the manufacturer markets would be likely to cause or contribute to a death or serious injury if the malfunction were to recur. Such reports are referred to as "initial reports." Manufacturers who subsequently obtain information about the event that was not known or was not available when the initial report was submitted, but which would have been required to be submitted as part of the initial report had that additional information been known or available, must file a supplemental report or "supplemental MDR" with FDA within thirty (30) days of receiving the additional information.

15. MDRs are one of the post-market surveillance tools FDA uses to monitor device performance, detect potential device-related safety issues, and contribute to benefit-risk assessments of devices.

16. A device is deemed to be "misbranded" under 21 U.S.C. § 352(t)(2) if a manufacturer fails or refuses to furnish any material information required by 21 U.S.C. § 360i respecting the device, including MDRs and supplemental MDRs. The FDCA prohibits the introduction of misbranded medical devices into interstate commerce, pursuant to 21 U.S.C. § 331(a).

**The Q180V and OMSC's MDR Reporting**

17. As the manufacturer of the Q180V, OMSC bore ultimate responsibility for the filing of MDRs to FDA for adverse events involving the Q180V anywhere in the world. Prior to April 2012, OAI personnel filed MDRs for OMSC for adverse events occurring anywhere in the world. In early 2012, OMSC shifted responsibility for preparing and filing MDRs for adverse events occurring outside of North and South America from OAI personnel in the United States to OMSC personnel in Japan.

18. OMSC employees received minimal training to prepare for this transfer of responsibilities, which left them uncertain about what information must be included in an MDR and in what circumstances a supplemental MDR must be filed. Some OMSC employees believed that they had inadequate resources to take over the responsibility, and informed supervisors that they needed additional training and resources to meet the MDR reporting requirements. Their requests for assistance were denied by OMSC management.

19. As detailed below, OMSC's lack of MDR training and inadequate resources contributed to OMSC's failure to file MDRs and supplemental MDRs between August 2012 and October 2014.

20. Between August 2012 and October 2014, sales of the OMSC-manufactured Q180V in the United States generated

approximately \$40 million in revenue and approximately \$33 million in total gross profit.

**Erasmus Medical Center (Netherlands) Infections**

21. Between January and April 2012, approximately 22 patients at the Erasmus Medical Center in the Netherlands were infected with *Pseudomonas aeruginosa* after the same Q180V duodenoscope was used on them. The same bacteria was detected in a sample collected from the device.

22. In March 2012, officials from Erasmus Medical Center notified an Olympus Corporation subsidiary in the Netherlands of the infections. In April 2012, employees of OMSC learned of the infections.

23. On or about May 25, 2012, OMSC filed an MDR concerning the infections at Erasmus Medical Center. The MDR stated, "OMSC can not [sic] conclusively determine the cause [sic] this event. However, it can be considered as a possible cause of this phenomenon that the patient infected from other than the endoscope and procedure such as environmental factor in the facility."

24. As OMSC was preparing the MDR, an independent expert, Dr. Arjo Loeve of Delft University of Technology in the Netherlands, disassembled the Erasmus duodenoscope in the presence of representatives from Olympus Europe and Erasmus. He took samples from various points on the scope, analyzed the

scope itself, and prepared a detailed report -- "Investigation Report on Scope G-206" -- commonly referred to as "the Delft Report."

25. The Olympus subsidiary in the Netherlands received a draft of the Delft Report in Dutch on or about May 22, 2012, and a final version of the Delft Report in Dutch on or about June 30, 2012. OMSC received an English translation of the Delft Report on or about August 6, 2012.

26. The Delft Report stated that *Pseudomonas aeruginosa* was found on the cap of the scope and that brownish deposits were found at several places, including in the closed elevator channel (referred to in the Delft Report as the propulsion cavity). The Delft Report observed that the tip of the scope had cracks, corners, and cavities that were very difficult to reach with a brush. The Delft Report further stated that the O-ring -- which was designed to seal the closed elevator wire channel -- likely failed to function properly and that it was "likely that moisture and/or biological material from the shaft or the tip of the endoscope entered the propulsion cavity and has remained and/or grown there." The Delft Report's conclusions included that the scope's tip had various cracks, corners, and crevices that could harbor bacteria and could be cleaned only with great difficulty; that deposits were found at various places, including in an area that should

have been sealed from liquids; and that the O-ring did not guarantee a reliable seal. The Delft Report recommended immediate further investigation of all such scopes, updating the cleaning instructions, and improving the quality of the seals.

27. OMSC was required to supplement the initial MDR regarding the Erasmus adverse events upon receiving the Delft Report, but did not do so.

28. FDA did not learn of the Delft Report and its findings and recommendations until October 2014. Upon learning of the Delft Report independently, FDA contacted OAI, asked if Olympus was aware of the report, and encouraged Olympus to obtain and read the report -- a report that, unbeknownst to FDA, OMSC had received more than two years earlier. FDA communicated concerns regarding information included in the Delft Report and asked OAI for additional information.

29. After evaluating information regarding the infections at Erasmus, including the Delft Report, and learning about the infections at Clinique de Bercy described below, the Dutch Health Inspectorate ("IGZ") -- a Dutch government agency akin to FDA -- in December 2012 met with employees of Olympus Corporation subsidiaries in Europe, and Olympus Corporation subsidiaries in Europe agreed to send a Field Safety Corrective Action ("FSCA") to customers in the Netherlands. A subsidiary of Olympus Corporation in Europe prepared the notice and sent it

to all European customers in or around January 2013. The January 2013 FSCA reminded customers to pay close attention to the Q180V's reprocessing instructions. An accompanying "Quick Reference Guide" suggested use of a small brush -- the MAJ-1888, which was an optional accessory available only in Europe -- to obtain deeper access to the forceps elevator during reprocessing.

30. On or about March 13, 2015, over two years and seven months after receiving an English translation of the final Delft Report, OMSC filed supplemental MDRs concerning each of the 22 Erasmus patients who were infected with *Pseudomonas aeruginosa* after the same Olympus TJF-Q180V duodenoscope was used on them. The supplemental MDRs stated that Delft University had disassembled the duodenoscope and found brownish deposits on both sides of the O-ring.

**Clinique de Bercy (France) Infections**

31. In or around November 2012, three patients at Clinique de Bercy in France were infected with *Escherichia coli* after the same Olympus TJF-Q180V duodenoscope was used on them.

32. In November 2012, officials from Clinique de Bercy notified an Olympus Corporation subsidiary in France of the infections and on or about November 29, 2012, employees of OMSC learned of the infections.

33. On or about December 20, 2012, OMSC filed MDRs concerning each of the three patients who were infected at Clinique de Bercy. The MDRs stated that the subject device "will be sent to an independent microbiology laboratory for microbiological testing. At the present time, the exact cause of the reported phenomenon cannot be determined, however insufficient reprocessing and user handling cannot be ruled out as contributory factors. If significant additional information is received, a supplemental report will follow."

34. In or around November 2012, the French Agence Nationale de Sécurité du Médicament et des Produits de Santé ("ANSM") -- a French government agency akin to FDA -- started an inquiry into the infections at Clinique de Bercy. As part of its inquiry, ANSM directed Biotech Germande, an independent microbiological laboratory, to examine the cleanability of the Q180V duodenoscope used on the three infected patients at Clinique de Bercy.

35. On or about April 13, 2013, OMSC received a report prepared by Biotech Germande of the results of testing the Q180V used on the three infected patients at Clinique de Bercy. The Biotech Germande report stated that the Q180V was contaminated with various bacteria and that contamination remained after the duodenoscope was reprocessed according to OMSC's reprocessing instructions. The Biotech Germande report

concluded that "after completing a full cleaning/disinfection procedure according to the ministerial and endoscope manufacturer's guidelines, there is a risk of persistence of contamination . . ."

36. OMSC was required to supplement the initial MDRs regarding the Bercy adverse events upon receiving the Biotech Germande test results, but did not do so.

37. In or around June 2013, employees of Olympus Corporation subsidiaries in Europe met with representatives of ANSM to discuss the Q180V and the Biotech Germande test results. Employees of Olympus Corporation subsidiaries in Europe presented the use of the MAJ-1888 brush as a solution to ANSM's concerns regarding the cleanability of the Q180V. ANSM then required additional testing by Biotech Germande and another laboratory, Bonn University, regarding the cleanability of the Q180V using the MAJ-1888 brush.

38. In February 2014, Olympus Corporation subsidiaries in Europe shared the results of the additional testing of the Q180V with ANSM. These results included a finding by Biotech Germande that the MAJ-1888 improved the cleaning of the distal end of the Q180V, but the Q180V still had detectable residual contamination after reprocessing. Considering those results, ANSM informed Olympus on or about March 24, 2014, that "there remains a legitimate doubt

concerning [the cleanability of the Q180V] under real conditions of use." ANSM subsequently directed Olympus to amend the instructions for use in Europe to recommend use of the MAJ-1888 brush, and to prepare an FSCA to customers announcing the change.

39. On or about July 28, 2014, Olympus Corporation subsidiaries in Europe distributed the FSCA recommending use of the MAJ-1888 brush to European customers.

40. OMSC never filed supplemental MDRs concerning the Bercy adverse events with the results of the Biotech Germande testing, although in a table attached to an October 31, 2013, letter to FDA, Olympus mentioned that an independent laboratory detected environmental bacteria on the scope used at Bercy and that subsequent testing by the laboratory after reprocessing indicated debris on the distal end of the scope.

#### **Kremlin Bicetre (France) Infections**

41. On or about July 4, 2012, a subsidiary of Olympus Corporation in France received a report of five patients at Kremlin Bicetre in France who were infected with *Pseudomonas aeruginosa* after the same Olympus TJF-Q180V duodenoscope was used on them. On or about July 10, 2013, OMSC received an email from its subsidiary in Europe, which included a fax communication from ANSM referencing contamination of a scope at Kremlin Bicetre.

42. OMSC was required to file MDRs concerning the Kremlin Bicetre infections. OMSC did not file MDRs concerning the Kremlin Bicetre infections until on or about July 7, 2016.

**Later FDA Actions**

43. On or about September 17, 2013, as part of an FDA effort to review information regarding risks associated with the transmission of infections by all major duodenoscopes being marketed in the United States, FDA sent OMSC an additional information request ("AIR"), seeking more information about certain MDRs OMSC had filed relating to the Q180V and other scopes. Communications between FDA and Olympus ensued and FDA's review continued.

44. On or about February 19, 2015, FDA issued a Safety Communication -- Design of Endoscopic Retrograde Cholangiopancreatography (ERCP) Duodenoscopes May Impede Effective Cleaning -- informing users that the complex design of duodenoscopes may impede effective cleaning. The Safety Communication stemmed from the FDA effort's to assess the safety of all duodenoscopes - not just Olympus's Q180V duodenoscope - and applied to all duodenoscopes. The FDA Safety Communication noted that the design of duodenoscopes "causes challenges for cleaning and high-level disinfection . . . part of the scopes may be extremely difficult to access and effective cleaning of all areas of the duodenoscope may not be possible." The FDA

Safety Communication recommended that users “[f]ollow closely all manufacturer instructions for cleaning and processing,” and noted that “[e]ven though duodenoscopes are inherently difficult to reprocess, strict adherence to the manufacturer’s reprocessing instructions will minimize the risk of infection.”

45. On or about March 4, 2015, FDA released Updated Information for Healthcare Providers Regarding Duodenoscopes. FDA recommended that healthcare professionals inform patients of the potential risks of infection accompanying the use of duodenoscopes, and stated that FDA would continue evaluating “alternative cleaning protocols . . . and explore additional strategies to reduce the risk of infections. . . .” The FDA Updated Information noted that “FDA’s analysis indicates that the reported duodenoscope-associated infections have occurred in patients who have had procedures with duodenoscopes from all three manufacturers.” FDA noted that it was “not recommending that healthcare providers cancel ERCP procedures for their patients who need them.”

46. On or about March 26, 2015, FDA issued a Safety Communication -- New Reprocessing Instructions Validated for Model TJF-Q180V Duodenoscopes -- which announced new and validated reprocessing instructions for the Q180V. The new instructions included changes to processing procedures on

precleaning, manual cleaning, and manual high-level disinfection and required use of the MAJ-1888 brush.

---

Count One

**(Introduction of Misbranded Medical Devices  
into Interstate Commerce, 21 U.S.C. §§ 331(a) and 333(a)(1))**

47. The allegations contained in paragraphs 1 through 46 are realleged and incorporated herein as if set forth in full.

48. On or about April 10, 2014, in the District of New Jersey and elsewhere, defendant

**OLYMPUS MEDICAL SYSTEMS CORPORATION**

did introduce and deliver for introduction, and cause the introduction or delivery for introduction, into interstate commerce, misbranded (pursuant to 21 U.S.C. § 352(t)(2)) medical devices, including a device shipped to a hospital in New Jersey, which were misbranded due to OMSC's failure to file with FDA supplemental MDRs relating to infections at Erasmus Medical Center.

All in violation of 21 U.S.C. §§ 331(a) and 333(a)(1).

Count Two

**(Introduction of Misbranded Medical Devices  
into Interstate Commerce, 21 U.S.C. §§ 331(a) and 333(a)(1))**

49. The allegations contained in paragraphs 1 through 46 are realleged and incorporated herein as if set forth in full.

50. On or about June 19, 2014, in the District of New Jersey and elsewhere, defendant

**OLYMPUS MEDICAL SYSTEMS CORPORATION**

did introduce and deliver for introduction, and cause the introduction or delivery for introduction, into interstate commerce, misbranded (pursuant to 21 U.S.C. § 352(t)(2)) medical devices, including a device shipped to a hospital in New Jersey, which were misbranded due to OMSC's failure to file with FDA supplemental MDRs relating to infections at Clinique de Bercy.

All in violation of 21 U.S.C. §§ 331(a) and 333(a)(1).

**Count Three**

**(Introduction of Misbranded Medical Devices  
into Interstate Commerce, 21 U.S.C. §§ 331(a) and 333(a)(1))**

51. The allegations contained in paragraphs 1 through 46 are realleged and incorporated herein as if set forth in full.

52. On or about July 22, 2014, in the District of New Jersey and elsewhere, defendant

**OLYMPUS MEDICAL SYSTEMS CORPORATION**

did introduce and deliver for introduction, and cause the introduction or delivery for introduction, into interstate commerce, misbranded (pursuant to 21 U.S.C. § 352(t)(2)) medical devices, including a device shipped to a hospital in New Jersey, which were misbranded due to OMSC's failure to file with FDA MDRs relating to infections at Kremlin Bicetre.

All in violation of 21 U.S.C. §§ 331(a) and 333(a)(1).

**FORFEITURE ALLEGATION**

1. The allegations contained in all paragraphs of this Information are hereby re-alleged and incorporated by reference for the purpose of noticing forfeiture pursuant to Title 21, United States Code, Section 334 and Title 28, United States Code, Section 2461.

2. The United States hereby gives notice to the defendant that, upon conviction of the offenses charged in this Information, the government will seek forfeiture, in accordance with Title 21, United States Code, Section 334 and Title 28, United States Code, Section 2461(c), of any and all medical devices that were introduced into interstate commerce, contrary to the provisions of Title 21, United States Code, Section 331.

**Substitute Assets Provision**

3. If any of the property described above, as a result of any act or omission of the defendant:

- (a) cannot be located upon the exercise of due diligence;
- (b) has been transferred or sold to, or deposited with, a third party;
- (c) has been placed beyond the jurisdiction of the court;

(d) has been substantially diminished in value; or

(e) has been commingled with other property which cannot be divided without difficulty,

the United States of America will be entitled to forfeiture of substitute property up to the value of \$5 million for the property described in the forfeiture allegations set forth above, pursuant to Title 21, United States Code, Section 853(p), as incorporated by Title 28, United States Code, Section 2461(c).

Rachael A. Honig  
RACHAEL A. HONIG  
ATTORNEY FOR THE  
UNITED STATES ACTING UNDER  
AUTHORITY CONFERRED BY  
28 U.S.C. § 515

CASE NUMBER: 18-

---

---

**United States District Court  
District of New Jersey**

---

---

**UNITED STATES OF AMERICA**

**v.**

**OLYMPUS MEDICAL SYSTEMS CORPORATION**

---

---

**INFORMATION FOR**  
21 U.S.C. §§ 331(a) and 333(a)(1)

---

---

**RACHAEL A. HONIG**  
*ATTORNEY FOR THE UNITED STATES ACTING UNDER  
AUTHORITY CONFERRED BY 28 U.S.C. § 515  
NEWARK, NEW JERSEY*

---

---

JACOB T. ELBERG  
R. DAVID WALK, JR.  
ASSISTANT U.S. ATTORNEYS  
(973) 645-3987

---

---

USA-48AD 8  
(Ed. 1/97)